

A new liquid chromatography method for the simultaneous and sensitive quantification of lactose and lactulose in milk

Regina SCHUSTER-WOLFF-BÜHRING*, Ronnie MICHEL, Jörg HINRICHS

Department of Dairy Science and Technology, Institute of Food Science and Biotechnology,
University of Hohenheim, Stuttgart, Germany

Received 13 January 2010 – Revised 16 April 2010 – Accepted 30 April 2010

Published online 2 August 2010

Abstract – The disaccharide lactulose is known to have a large influence on human digestion, the effects of which range from prebiotic to laxative action. Lactulose can be derived from lactose by either alkaline isomerization or enzymatic transgalactosylation of fructose. The aim of this study was to establish a new analytical method for the quantification of both disaccharides lactose and lactulose using high performance liquid chromatography (HPLC) and evaporative light scattering detection (ELSD). The new method needs to provide good chromatographic separation of lactose and lactulose while being fast and reliable. Therefore, four analytical methods using different HPLC columns were developed and compared. A HPLC column with an amino-bonded polymeric matrix yielded better results compared to amino-bonded silica-phase resin or cation-exchange resin. The new method had detection limits for lactose and lactulose of 3.8 and 2.5 mg·L⁻¹ and was successfully applied to the quantification of lactulose and lactose in commercial milk products.

HPLC / lactulose / lactose / evaporative light scattering detection / method development

摘要 – 一种新的液相色谱法同时检测乳中乳糖和乳酮糖。乳酮糖为二糖，具有益生元的功能。乳酮糖除对人消化能力有较大的影响外，还具有对人体放松的作用。乳酮糖的制备可以采用乳糖在碱性条件下的异构化作用或者在酶催化下发生果糖的转半乳糖苷作用。本文建立了配以蒸发光散射检测器 (ELSD) 的高效液相色谱法 (HPLC) 测定乳糖和乳酮糖的定量检测方法。新的色谱条件必须保证乳糖和乳酮糖有较高的分离度，并且具有快速和准确的特点。因此，选用不同的色谱柱用四种分析方法对测定结果进行了比较和分析。与氨基键合硅胶柱或阳离子交换树脂柱相比，氨基键合柱可以将乳糖和乳酮糖较好地分离。所建立的方法对乳糖和乳酮糖的检测限分别为 3.8 和 2.5 mg·L⁻¹。该方法可以成功地用于乳制品中乳糖和乳酮糖的定量测定。

HPLC / 乳酮糖 / 乳糖 / 蒸发光散射检测 / 方法的建立

1. INTRODUCTION

The prebiotic disaccharide lactulose has been subject to clinical trials for 50 years

and is nowadays regarded as a valuable substance for both pharmaceutical and food applications [3, 4, 6, 17, 27, 28, 29]. Lactulose (4-O-β-D-galactopyranosyl-D-fructose)

*Corresponding author (通讯作者): gini@uni-hohenheim.de

can be synthesized by different reactions with the alkaline isomerization of lactose via *Lobry de Bruyn – Alberda van Ekenstein* – rearrangement being the only one applied in industrial scale [2]. An alternative way to produce lactulose is offered by a side reaction of lactose hydrolysis, the transgalactosylation of fructose [11, 13, 16]. Our current research focuses on this enzymatic transformation of milk-based lactose into lactulose and the precise determination of these sugars is a vital part of this work. Moreover, the precise analysis of lactose and lactulose in food or drug formulations is of special interest as both substances exert dose-dependent effects upon intake (lactose maldigestion, prebiotic or laxative action of lactulose).

The aim of the present work was to develop a sensitive yet fast and reliable method for the quantification of lactose and lactulose in milk-based products. A glance over available literature on lactulose demonstrates the broad diversity of applied analytical methods [14, 15, 26, 30]. Besides enzymatic assays lactulose is mainly analysed using chromatography, either thin-layer, liquid or gas chromatography. We decided to perform analysis via high performance liquid chromatography (HPLC) because it allows simultaneous and fast identification of mono-, di- and oligosaccharides. As for detection, we chose an evaporative light scattering detector (ELSD) which is commonly acknowledged for its sensitivity [7]. Our method development is based on the comparison of four different HPLC columns in regard to their separation performance of lactose and lactulose.

2. MATERIALS AND METHODS

Materials. Carbohydrate standards lactose, lactulose, glucose, galactose and fructose of analytical grade were obtained from Sigma-Aldrich Inc. (Munich, Germany) and Carl Roth GmbH & Co.

KG (Karlsruhe, Germany). Distilled water (conductivity $< 4 \mu\text{s}\cdot\text{cm}^{-1}$) was used for dilution. Commercially available lactose- and lactulose-containing products were: lactose-free yoghurt and sterilized milk (Breisgaumilch GmbH, Breisgau i.S., Germany). Skim milk was processed in the Dairy for Research and Training at the University of Hohenheim by separation of raw milk (62 °C) and pasteurization (72–75 °C, 30 s).

Sample preparation. For chromatographical analysis, commercial samples were prepared by dilution with water, protein precipitation by addition of Carrez-solution and twofold filtration. Every sample was dissolved with distilled water, while the dilution factor was varied from 1:10 to 1:50, depending on the sample. Before the dilution was made up to its intended volume, 0.5 mL of each Carrez-solution I and II were added sequentially and the mixture was left for 10 min, which was found to be sufficient for complete protein precipitation. Carrez-solution I and II were prepared according to Indyk et al. [10]. Afterwards, the diluted milk product samples were made up to full volume and then filtered through filter paper (Whatman's No. 1, $\text{Ø} = 150 \text{ mm}$; Whatman GmbH, Dassel, Germany). The resulting permeate was further passed through a membrane with a pore size of $0.20 \mu\text{m}$ (Chromafil RC-20/25, Macherey-Nagel GmbH & Co. KG, Düren, Germany). The extract was injected into the HPLC-system following a 1:2 dilution with acetonitrile (gradient grade, Merck KGaA, Darmstadt, Germany).

HPLC-system. Sugar separation was carried out using a chromatographic unit from Agilent Technologies Inc. (Santa Clara, CA, USA). The unit (Series Agilent 1100) consisted of a vacuum degasser, a quaternary pump, an autosampler and a column oven. Sugar detection resulted from evaporative light scattering (Alltech ELSD 2000 ES, Alltech Associates Inc., Deerfield, IL, USA).

Table I. Specification of employed HPLC columns.

Trade name and manufacturer	Column dimension in mm	Packaging properties ^a	Separation mode
Rezex RCM-Monosaccharide Ca ⁺ , Phenomenex Inc., Torrance, CA, USA	300 × 7.8	Sulfonated crosslinked styrene divinylbenzene copolymer with calcium couterions	Ion exclusion
Prevail Carbohydrate ES, Alltech Associates Inc., Deerfield, IL, USA	250 × 4.6	Five micrometer spherical polymer beads coated with amino-based proprietary bonding material	Hydrogen bonding
Sphere Clone NH ₂ , Phenomenex Inc., Torrance, CA, USA	250 × 4.6	Five micrometer Silica particels bonded with a monolayer of 3-aminopropylsilan	Hydrogen bonding
Zorbax Carbohydrate Analysis, Agilent Technologies Inc., Santa Clara, CA, USA	250 × 4.6	Five micrometer “Zorbax – silica” particels bonded with a monolayer of 3-aminopropylsilan	Hydrogen bonding

^a According to supplier.

Stationary phases. Four HPLC Columns with different stationary phases were tested. A detailed description of the employed columns is given in Table I.

One hundred per cent distilled water was used as mobile phase for analysis with column Rezex RCM-Monosaccharide Ca⁺, while the other columns were eluted with a mixture of distilled water and acetonitrile (CH₃CN). Composition of mobile phases and further analysis conditions such as flow rate and column temperature were optimized during this work. Original detector settings were: no Impactor, Gain 1, 80 °C tube temperature and 2.5 L·min⁻¹ nitrogen flow.

Method development. Column performances were optimized with regard to retention time (t_R) and peak resolution (R_S):

$$R_S = \frac{2(t_{R2} - t_{R1})}{w_{b1} + w_{b2}},$$

where w_{b1} is the peak width for substance 1 in minutes and w_{b2} is the peak width for substance 2 in minutes.

According to Ruecker et al. [25] a $R_S > 1.4$ indicates the baseline separation of two substances.

Baseline noise (N). The value was calculated based on a sixfold measurement of maximum baseline amplitude by injection of 20 µL acetonitrile:

$$N = f_{\max} - f_{\min},$$

where f_{\max} is the maximum baseline swing in mAU and f_{\min} is the minimum baseline swing in mAU.

Limit of quantification (LOQ) and limit of detection (LOD). The values for LOQ and LOD with regard to lactose and lactulose were calculated based on the ratio of peak height (S) to the baseline noise (N) and the calibration curves for both disaccharides. The limit of sugar quantification

(LOQ) is defined as $S/N = 10$, while the limit of sugar detection (LOD) is $S/N = 3$.

Recovery. The recovery of lactulose was measured by adding 0.4 g lactulose to 10 mL skim milk ($c = 40 \text{ mg}\cdot\text{mL}^{-1}$). After sample preparation, milk was analysed along with a lactulose solution with the same concentration. Recovery of lactulose is expressed as the peak area of lactulose in the milk sample related to the peak area of lactulose solution.

Repeatability. The repeatability of peak areas of lactose and lactulose was determined with six consecutive chromatographic analyses of solutions with analyte concentration of $0.01 \text{ g}\cdot\text{L}^{-1}$.

Quantification. Two six-point calibration curves were made for sugar standards lactose, lactulose, glucose, galactose and fructose, ranging from 0.005 to $0.04 \text{ g}\cdot\text{L}^{-1}$ and from 0.04 to $0.3 \text{ g}\cdot\text{L}^{-1}$. Carbohydrates were identified by comparing retention times with sugar standards. Peak areas of sugar standards were used for quantification.

3. RESULTS

3.1. Comparison of stationary phases

The objective of this work was to determine among four different columns the column with the best separation of lactose, lactulose, glucose, galactose and fructose, with our main focus being the separation and further quantification of the disaccharides. Therefore, chromatographic settings such as composition and flow rate of the mobile phase and column temperature were varied from the original recommendations stated by producer. Retention time (t_R) and peak resolution (R_S) were calculated in order to compare the separation performance of the different stationary phases. While retention time should be short resulting in a fast analysis, R_S should be higher than 1.4. At this value, it is defined that

the mirrored substance is clearly separated up to the baseline from the previous peak [25]. Table II reflects a few steps of method optimization for the chosen columns, with the best settings for each column indicated in bold symbols.

Table II makes clear that column Rezex differs principally from the three other employed columns as it elutes the disaccharides lactose and lactulose prior to the monosaccharides, which can be explained by the size exclusion effect of cationic resins [21]. The best peak resolution for lactose and lactulose was achieved at a $0.6 \text{ mL}\cdot\text{min}^{-1}$ flow rate of distilled water at $80 \text{ }^\circ\text{C}$, when lactose and lactulose are clearly separated ($R_S = 1.46$). Lactulose could not be fully separated from the following monosaccharide glucose, which elutes only 30 s after lactulose, resulting in a R_S of 0.59. The separation performance of column Rezex at optimized chromatographic conditions is depicted in Figure 1.

Columns Zorbax and SphereClone generally showed the same elution order for the standard sugar solution, eluting the monosaccharides fructose, galactose and glucose prior to the disaccharides lactulose and lactose. Retention times of lactulose and lactose with column Zorbax under optimized conditions were 15.2 and 16.2 min and the subsequent R_S for these peaks 1.03, thus yielding in an insufficient separation. Disaccharide separation with column SphereClone was achieved with a broad separation of lactulose and lactose ($R_S = 2.02$) and retention times of 20.4 and 22.6 min, respectively. Chromatograms with both columns are depicted in Figure 1.

Separation performance of column Prevail is represented in Table I: monosaccharides were eluted prior to lactulose and lactose and the difference to columns SphereClone and Zorbax is the altered elution of galactose and glucose. Optimization of column settings leads to a clear separation of lactose and lactulose up to the baseline ($R_S = 5.68$) at relatively short retention

Table II. Retention times (t_R) and resolution factor (R_S) of sugar standards for different chromatographic settings and columns Rezex RCM Monosaccharide, Zorbax Carbohydrate, SphereClone NH₂ and Prevail Carbohydrate ES. Bold: best chromatographic performance regarding both t_R and R_S for lactose and lactulose.

Column	Column settings			Carbohydrates in order of elution									
	Mobile phase %ACN/%H ₂ O	Flow rate in mL·min ⁻¹	Column temperature in °C	(t_R) in min	(R_S)	(t_R) in min	(R_S)	(t_R) in min	(R_S)	(t_R) in min	(R_S)	(t_R) in min	(R_S)
Rezex	0/100	0.7	80	Lactose 9.13	–	Lactulose 9.88	1.36	Glucose 10.3	0.30	Galactose 11.9	1.41	Fructose 12.8	2.31
	0/100	0.6	80	9.56	–	10.4	1.41	10.9	0.59	12.1	1.51	13.8	2.57
Zorbax	75/25	1.4	50	Fructose 4.65	–	Galactose 5.27	1.03	Glucose ^a		Lactulose 8.65	4.82	Lactose 9.27	0.67
	80/20	1.4	40	6.37	–	7.63	1.73	^a		16.4	7.80	18.0	0.98
	75/25	1.0	30	7.71	–	8.60	1.49	9.26	0.80	15.2	6.23	16.2	1.03
SphereClone	60/40	0.6	25	6.24	–	6.61	1.13	6.90	0.63	8.42	3.02	8.81	0.82
	70/30	0.8	25	8.96	–	10.1	2.41	10.7	0.94	16.1	6.61	17.6	1.79
	73/27	0.8	25	10.2	–	11.7	2.89	12.6	1.17	20.4	8.13	22.6	2.02
Prevail Carbohydrate	75/25	0.7	30	Fructose 10.6	–	Glucose 13.0	6.77	Galactose 13.8	1.92	Lactulose 17.8	8.07	Lactose 22.3	7.33
	75/25	0.9	30	8.23	–	10.1	6.32	10.8	1.89	13.8	7.37	17.4	6.81
	70/30	0.9	30	6.36	–	7.38	4.45	7.77	1.52	8.83	3.65	10.5	5.03
	70/30	0.9	25	6.96	–	8.24	5.12	8.70	1.64	10.3	4.91	12.5	5.68

^a Substance eluted simultaneously with previous analyte; ACN acetonitrile; t_R , retention time; R_S , peak resolution.




Figure 1. Chromatogram of standard solution with lactose, lactulose, glucose, galactose and fructose (respective concentration = $0.1 \text{ g}\cdot\text{L}^{-1}$) with columns Rezex RCM Monosaccharide (1a), Zorbax Carbohydrate (1b), SphereClone NH_2 (1c) and Prevail Carbohydrate ES (1d) under optimized chromatographic settings (see Tab. II) with $20 \mu\text{L}$ injection. ELSD-settings: tube temperature $85 \text{ }^\circ\text{C}$; nitrogen flow rate $2.5 \text{ L}\cdot\text{min}^{-1}$.

times of 10.3 and 12.6 min. While the difference in optimized retention times of lactose and lactulose is nearly the same as with column SphereClone (~ 2.4 min), column Prevail shows a better peak resolution ($R_S = 5.68$ compared to 2.02). This is mainly due to the smaller width of the peaks for lactose and lactulose obtained with column Prevail, shown in Figure 1. Generally, chromatographic analysis with column Prevail had the best accordance with our requirements: a clear peak resolution for lactose and lactulose ($R_S > 1.4$) and a short analysis time.

3.2. Method evaluation

Comparing four different stationary phases in regard to the separation of lactose and lactulose, we found that column Prevail featured the best peak resolution at a short analysis time. Optimized chromatographic settings were: isocratic elution with a mixture of acetonitrile and distilled water

(70:30%), a mobile phase flow rate of $0.9 \text{ mL}\cdot\text{min}^{-1}$ at $25 \text{ }^\circ\text{C}$. The new method was then characterized by determining the average noise, the LOD and quantification and the recovery of lactulose.

The baseline noise is influenced by both column resin and detector settings. In order to characterize the resin-associated noise of the chosen method, we compared column Prevail to columns Zorbax and SphereClone with identical detector settings. The average noise, measured with the three columns, is displayed in Table III.

Table III clarifies that column Prevail exerts with 0.3 mAU considerably less baseline noise than columns Zorbax and SphereClone. Especially the latter has a high baseline noise, which can also be seen in Figure 1c. An intensive baseline noise of an amino-bonded resin can be caused by hydrolytic breakdown and subsequent detection of aminogroups [9]. This effect is enforced with increasing percentage of water in the mobile phase.

Table III. Detector noise for different columns. Column settings: see [Table II](#), indicated in bold letters; ELSD-settings: tube temperature 85 °C; nitrogen flow rate 2.5 L·min⁻¹.

Column	Zorbax Carbohydrate	SphereClone NH ₂	Prevail Carbohydrate
Noise in mAU ± SD	2.6 ± 0.8	34 ± 9.9	0.33 ± 0.09

$n = 6$; SD, standard deviation.

Table IV. Limit of detection (LOD) and quantification (LOQ), repeatability and recovery of HPLC-Column Prevail Carbohydrate ES. Column settings: see [Table II](#), indicated in bold letters; ELSD-settings: tube temperature 85 °C; nitrogen flow rate 2.5 L·min⁻¹.

	LOD in mg·L ⁻¹	LOQ in mg·L ⁻¹	Repeatability ^a		Recovery ^b
			Mean ± SD in mAU	Rel. SD in %	Mean ± SD in mAU
Lactulose	2.5	15.0	451 ± 17	3.8	99.5 ± 0.8
Lactose	3.8	17.3	496 ± 13	2.6	n.d.

n.d., not determined; ^a $n = 6$; ^b $n = 3$; SD, standard deviation.

The determination of column noise (N) allows the calculation of the LOD and the LOQ for lactose and lactulose. Besides, repeatability and recovery of the evaluated method were determined for lactose and lactulose, respectively. [Table IV](#) summarizes the evaluated parameters for the developed method with column Prevail. Both lactose and lactulose have low limits of detection with 3.8 and 2.5 mg·L⁻¹, respectively. Repeatability of lactose and lactulose detection was evaluated with relative standard deviations of 2.6% and 3.8%, while 99.5% of lactulose could be recovered after sample preparation.

3.3. Application to real samples

The developed method for simultaneous detection of lactose and lactulose was applied to the analysis of commercial samples which contain low amounts of lactose or lactulose. [Figure 2](#) shows chromatograms of sterilized milk (a) and lactose-reduced yoghurt (b). Both demonstrate the methods' performance in regard to the detection of small amounts of disaccharides in milk

matrices. The sample of sterilized milk contained (1.81 ± 0.07) g·L⁻¹ lactulose ($n = 6$), while residual lactose concentration in lactose-reduced yoghurt was (0.72 ± 0.09) g·kg⁻¹ ($n = 3$). Adachi [1] firstly reported the presence of lactulose in intensively heat-treated milk, where it was formed due to an isomerization of lactose. Since then, the presence of lactulose in heated milk has been used as a parameter to differentiate between UHT-treatment and sterilization of milk [14]. The calculated lactulose concentration of 1.81 g·L⁻¹ is relatively high for sterilized milk [6], but may be caused by an additional heating of milk before sterilization. Note that the sample displayed in [Figure 2a](#) was strongly diluted in order to quantify both disaccharides properly during one analytical run. This procedure triggered a lactulose peak which converges towards the LOQ of lactulose (see [Tab. IV](#)). Chromatographic analysis of lactose-reduced yoghurt shows that residual lactose concentration is < 0.1%. This concentration is acceptable for consumers with lactose maldigestion. The reduction of lactose by enzymatic hydrolysis leads to




Figure 2. HPLC-chromatogram of sterilized milk (2a) and lactose-reduced yoghurt (2b) with HPLC-Column Prevail Carbohydrate ES. Column settings: see [Table II](#), indicated in bold letters; ELSD-settings: tube temperature 85 °C; nitrogen flow rate 2.5 L·min⁻¹.

the release of glucose and galactose, as can be seen in [Figure 2b](#). Both monosaccharides appear in such high concentration that they coelute as one peak. It has to be noted that chromatographic separation of glucose and galactose deteriorates with increasing column usage ($R_S < 1$ after 150 runs). This effect is probably caused by chemical reactions between reducing sugars and the aminophase attached to the packaging material [10, 23].

4. DISCUSSION

4.1. Comparison of stationary phases

Martínez-Castro et al. [15] reviewed several chromatographical methods for analysis of lactulose without determining the analytical differences between different types of HPLC-columns. HPLC analysis of lactulose and/or lactose has been previously performed by cationic ion-exchange resins [8, 16, 22, 24, 25, 30] and amino-silica-based stationary phases [5, 19, 20, 32]. Paseephol et al. [20], who investigated the isomerization of lactose into lactulose, emphasized the difficulty to quantify lactulose in the presence of lactose. Facing the same problem, Kim et al. [11] analysed their complex reaction mixture twofold: fructose and lactulose with an amino-modified silica packed column, while galactose and lactose were measured with

a cation-exchange column. A comparison of a cation-exchange resin and an amino-modified phase in regard to the separation of lactose and lactulose has not been made yet. Nevertheless, comparative studies on the separation of mono- and disaccharides are provided by Indyk et al. [10] and Muir et al. [18]. Both authors found that amino-bonded columns are more suitable for a clear separation of di- and even oligosaccharides, while cation-exchange resins feature a better separation of monosaccharides. Considering the different elution order of these resins, it can be generally stated that early eluted compounds are likely to undergo coelution while longer retained substances become more separated from each other. Our data support this conclusion, as we achieved better peak separation for lactose and lactulose with the amino-based resin ([Tab. II](#)). The three columns with amino-bonded resins employed in this study differ in regard to the packaging material: while columns Zorbax and Sphere-Clone consist of a silica gel, column Prevail is packed with a polymeric gel. This column has been previously shown to exert efficient separation of mono- and disaccharides in wine within a single run [12].

4.2. Method development

Previous reports on evaporative light scattering detection of mono- and disaccharides

have already proven the great sensitivity of this detection mode. Wei and Ding [31] found a LOD for lactose of $60 \text{ mg}\cdot\text{L}^{-1}$, while La Pera et al. [12] reported a LOD of $30 \text{ mg}\cdot\text{L}^{-1}$ and a LOQ of $100 \text{ mg}\cdot\text{L}^{-1}$. On the contrary, Chávez-Servin et al. [5] who used refractive index detection achieved a LOD for lactose of $250 \text{ mg}\cdot\text{L}^{-1}$. In terms of repeatability and recovery of lactose and lactulose the comparison with published data on [5, 10, 12] additionally validates the suitability of the developed method for reliable and sensitive detection as well as for quantification of lactose and lactulose.

5. CONCLUSION

We developed a HPLC method for simultaneous detection and quantification of lactose and lactulose. Among four different HPLC columns a column with an amino-modified polymeric resin showed the best peak resolution (R_S) for the disaccharides lactose and lactulose ($R_S = 5.68$) within a short retention time. The developed HPLC method featured a high sensitivity ($\text{LOD} < 3 \text{ mg}\cdot\text{L}^{-1}$) and precision (recovery $> 99\%$). Analysis of lactulose-containing sterilized milk and lactose-reduced yoghurt confirmed the method's suitability for detection and quantification of lactulose and lactulose.

Acknowledgements: This Research Project was supported by FEI (Forschungskreis der Ernährungsindustrie e.V., Bonn), the AiF and the Ministry of Economics and Technology. Project No.: AiF 14787N.

REFERENCES

- [1] Adachi S., Formation of lactulose and tagatose from lactose in strongly heated milk, *Nature* 181 (1958) 840–841.
- [2] Aider M., De Halleux D., Isomerization of lactose and lactulose production: review, *Trends Food Sci. Technol.* 18 (2007) 356–364.
- [3] Ballongue J., Schumann C., Quignon P., Effects of lactulose and lactitol on colonic microflora and enzymatic activity, *Scand. J. Gastroenterol. Suppl.* 222 (1997) 41–44.
- [4] Bouhnik Y., Attar A., Joly F., Riottot M., Dyard F., Flourié B., Lactulose ingestion increases faecal bifidobacterial counts: a randomised double-blind study in healthy humans, *Eur. J. Clin. Nutr.* 58 (2004) 1658–1664.
- [5] Chávez-Servin J.L., Castellote A.I., López-Sabater M.C., Analysis of mono- and disaccharides in milk-based formulae by high-performance liquid chromatography with refractive index detection, *J. Chromatogr. A* 1043 (2004) 211–215.
- [6] Claeys W., Van Loey A., Hendrickx M., Formation kinetics of hydroxymethylfurfural, lactulose and furosine in milk heated under isothermal and non-isothermal conditions, *J. Dairy Res.* 70 (2003) 287–301.
- [7] Clement A., Yong D., Brechet C., Simultaneous identification of sugars by HPLC using evaporative light scattering detection (ELSD) and refractive index detection (RI). Application to plant tissues, *J. Liq. Chromatogr.* 15 (1992) 805–817.
- [8] Dendene K., Guihard L., Balanec B., Bariou B., Study of the separation of lactose, lactulose and galactose by liquid chromatography using cationic ion-exchange resin columns, *Chromatographia* 41 (1995) 561–567.
- [9] Herbreteau B., Vilette V., Lafosse M., Dreux M., Analysis of oligosaccharides using aminobonded silica gel and ternary eluent with evaporative light scattering detection, *Fresenius J. Anal. Chem.* 351 (1995) 246–250.
- [10] Indyk H.E., Edwards M.J., Wollard D.C., High performance liquid chromatographic analysis of lactose-hydrolysed milk, *Food Chem.* 57 (1996) 575–580.
- [11] Kim Y.S., Park C.S., Oh D.K., Lactulose production from lactose and fructose by a thermostable β -galactosidase from *Sulfolobus solfataricus*, *Enz. Microb. Technol.* 39 (2006) 903–908.
- [12] La Pera L., Di Bella G., Magnesi R., Lo Turco V., Dugo G., Analysis of carbohydrates in sicilian DOC Wines by high performance liquid chromatography with evaporative light scattering detection, *Ital. J. Food Sci.* 19 (2007) 319–328.
- [13] Lee Y., Kim C., Oh D.K., Lactulose production by β -galactosidase in permeabilized cells of *Kluyveromyces lactis*, *Appl. Microbiol. Biotechnol.* 64 (2004) 787–793.

- [14] Marconi E., Messia M.C., Amine A., Moscone D., Vernazza F., Stocchi F., Pallechi G., Heat-treated milk differentiation by a sensitive lactulose assay, *Food Chem.* 84 (2004) 447–450.
- [15] Martínez-Castro I., Calvo M.M., Olano A., Chromatographic determination of lactulose, *Chromatographia* 23 (1987) 132–136.
- [16] Mayer J., Conrad J., Klaiber I., Lutz-Wahl S., Beifuss U., Fischer L., Enzymatic production and complete Nuclear Magnetic Resonance assignment of the sugar lactulose, *J. Agric. Food Chem.* 52 (2004) 6983–6990.
- [17] Mizota T., Mori T., Yaeshima T., Yanagida T., Iwatsuki M., Ishibashi N., Tamura Y., Fukuwatari Y., Effects of low dosages of lactulose on the intestinal function of healthy adults, *Milchwissenschaft* 57 (2002) 312–315.
- [18] Muir J.G., Rose R., Quarania R., Liels K., Barrett J.S., Shepard S.J., Gibson P.R., Measurement of short-chain carbohydrates in common Australian vegetables and fruits by High-Performance Liquid Chromatography, *J. Agric. Food Chem.* 57 (2009) 554–565.
- [19] Nikolov Z.L., Meagher M.M., Reilly P.J., High performance liquid chromatography of disaccharides on amine bonded silica columns, *J. Chromatogr. A* 319 (1985) 51–57.
- [20] Paseephol T., Small D.M., Sherkat F., Lactulose production from milk concentration permeate using calcium carbonate-based catalysts, *Food Chem.* 111 (2008) 283–290.
- [21] Pedruzzi I., Borges da Silva E.A., Rodrigues A.E., Selection of resins, equilibrium and sorption kinetics of lactobionic acid, fructose, lactose and sorbitol, *Sep. Purif. Technol.* 63 (2008) 600–611.
- [22] Pedruzzi I., Malvessi E., Mata V.G., Silva E.A.B., Silveira M.M., Rodrigues A.E., Quantification of lactobionic acid and sorbitol from enzymatic reaction of fructose and lactose by high-performance liquid chromatography, *J. Chromatogr. A* 1145 (2007) 128–132.
- [23] Porsch B., High-performance liquid chromatography of sugar mixtures containing xylose and arabinose on primary amino-bonded phases, *J. Chromatogr.* 253 (1982) 49–54.
- [24] Reimerdes E.H., Rothkitt K.-D., Ionenchromatographische Bestimmung von Lactulose und Epilactose neben Lactose und anderen Kohlenhydraten in Milch und Milchprodukten [Analysis of lactulose and epilactose in the presence of lactose and other carbohydrates in milk and milk products by ion exchange chromatography], *Z. Lebensm. Unters. Forsch.* 181 (1985) 408–411.
- [25] Ruecker G., Neugebauer M., Willems G.G., *Instrumentelle Pharmazeutische Analytik [Instrumental pharmaceutical analysis]*, 1st edn., Wissenschaftliche Verlagsgemeinschaft Stuttgart, Stuttgart, Germany, 2001.
- [26] Ruiz-Matute A.I., Sanz M.L., Corzo N., Martín-Alvarez P.J., Ibáñez E., Martínez-Castro I., Olano A.J., Purification of lactulose from mixtures with lactose using pressurized liquid extraction with ethanol-water at different temperatures, *J. Agric. Food. Chem.* 55 (2007) 3346–3350.
- [27] Schuhmann C., Medical, nutritional and technological properties of lactulose. An update, *Eur. J. Nutr.* 41, Suppl. 1 (2002) 17–25.
- [28] Seki N., Hamano H., Iiyama Y., Asano Y., Kokubo S., Yamauchi K., Tamaura Y., Uenishi K., Kudou H., Effect of lactulose on calcium and magnesium absorption: A study using stable isotopes in adult men, *J. Nutr. Sci. Vitaminol.* 53 (2007) 5–12.
- [29] Tuohy K., Ziemer C., Klinder A., Knoebel Y., Pool-Zobel B., Gibson G., Human volunteer study to determine the prebiotic effects of lactulose powder on human colonic microbiota, *Microb. Ecol. Health Dis.* 14 (2002) 165–173.
- [30] Verhaar L.A.T., Van der Aalst M.J.M., Beenackers J.A.W.M., Kuster B.F.M., Liquid chromatography of sugars on silica-based stationary phases, *J. Chromatogr.* 170 (1979) 313–328.
- [31] Wei Y., Ding M.-Y., Analysis of carbohydrates in drinks by high-performance liquid chromatography with a dynamically modified amino column and evaporative light scattering detection, *J. Chromatogr. A* 904 (2000) 113–117.
- [32] Zokae F., Kaghazchi T., Zare A., Soleimani M., Isomerization of lactose to lactulose-study and comparison of three catalytic systems, *Process Biochem.* 37 (2002) 629–635.